

**Product Name :**  
Expansion Index Test Apparatus**Product Code :**  
TVETSO30023**Description :**

Expansion Index Test Apparatus

**Technical Specification :**

Base and Collar: Cast aluminum construction; brass hold-down rods.

Construction: The unit consists of a cast aluminum base and collar, stainless steel specimen ring.

Specimen Ring: Stainless steel; 101.6 x 25.4 mm (4" internal diam. x 1" h.).

Loading Weight: Brass; 5.7 kg (12.6 lbs.).

Dial Indicator: 0.3" range x 0.0001" graduations. 8 mm range x 0.002 mm graduations.

Test Duration: 24 hours or less on the cured sample.

Porous Stones: 101.4 x 12.7 mm (4" diam. x 1/2" thick).



## Equipments Exporters

**Website:** [www.equipmentsexporters.com](http://www.equipmentsexporters.com), **Email:** [sales@equipmentsexporters.com](mailto:sales@equipmentsexporters.com)

**Address:** 75, Lajpat Nagar-IV, New Delhi-110024 **Phone:** +91-9311469084