

Product Name :
Cross Flow Heat Exchangers**Product Code :**
ATHE0029**Description :**

Cross Flow Heat Exchangers

Technical Specification :

- A variable slide valve controls the air flow.
- These enable students to measure the static pressure difference across the rods.
- The copper element is of known thermal capacity and includes a built-in thermocouple.
- The equipment allows students to quickly assess heat transfer rates by forced convection.
- They monitor the rate of cooling of a body of known thermal capacity in an air flow.
- It is a horizontal wind tunnel with a contraction cone, a working section, a diffuser, a constant speed fan, and an exhaust with silencer.
- A second thermocouple at the inlet to the wind tunnel measures the temperature of air entering the heat exchanger.
- Pitot traverse can measure air velocity at any vertical point in the working section, either before or after the rods.
- The working section includes a series of rods arranged in a matrix and at right-angles to the direction of air flow. To do experiments, students can remove any one of these rods and replace it with a cylindrical copper

element.

- The base of the working section includes two static pressure tapings: one before the rods and one afterwards.



Equipments Exporters

Website: www.equipmentsexporters.com, **Email:** sales@equipmentsexporters.com

Address: 75, Lajpat Nagar-IV, New Delhi-110024 **Phone:** +91-9311469084