

**Product Name :**  
Compaction Factor Apparatus**Product Code :**  
ALABS-A19-159**Description :****Compaction Factor Apparatus**

The apparatus is used for determining the workability of fresh concrete, provided the maximum size of the aggregate does not exceed 38mm. The test is particularly useful for concrete mixes of very low workability where true slump values are not reliable.

**Specifications :**

It consists of two rigid conical hoppers and a cylinder mounted on a rigid metal frame. The lower openings of the hoppers are fitted with hinged trap-doors having quick release catches. A circular metal plate is provided to cover the top of the cylinder. Supplied complete with one plasterer's trowel and one tamping rod, 16mm dia x 600mm long, one end rounded.

**Address:** 75, Lajpat Nagar-IV, New Delhi-110024 **Phone:** +91-9311469084

